

CITY OF LOS ANGELES

CALIFORNIA



ERIC GARCETTI
MAYOR

BOARD OF
BUILDING AND SAFETY
COMMISSIONERS

HELENA JUBANY
PRESIDENT

VAN AMBATIELOS
VICE-PRESIDENT

E. FELICIA BRANNON
VICTOR H. CUEVAS
SEPAND SAMZADEH

DEPARTMENT OF
BUILDING AND SAFETY
201 NORTH FIGUEROA STREET
LOS ANGELES, CA 90012

RAYMOND S. CHAN, C.E., S.E.
SUPERINTENDENT OF BUILDING
INTERIM GENERAL MANAGER

October 17, 2013

Attn: Mairi Miller
Nanometrics, Inc.
250 Herzberg Rd.
Kanata, Ontario, Canada K2K2A1

Electrical Testing Laboratory

2319 Dorris Place, Los Angeles, CA 90031
(323) 224-2168

www.LADBS.org

Approved:	September 14, 2013
Expiration:	September 14, 2014
Application No:	500465
Name of Item:	Strong Motion Accelerograph
Manufacturer:	Nanometrics, Inc.
Model:	TITAN EA, TITAN SMA

Conditions of Approval:

1. Not Suitable for Hazardous / Classified Areas.
2. This device shall only be used and maintained by qualified and trained personnel.
3. This equipment shall only be used in dry locations.
4. This device is only evaluated for fire and shock hazard, its performance and accuracy have not been investigated.

The above mentioned item has been examined and found to be in substantial compliance with the standards established for such equipment. This approval shall expire on the date indicated unless revoked sooner for cause.

Each approved item shall bear the approval label of the Department, attached in a manner determined by the standards, or otherwise made readily visible for inspection without disassembly. Approval labels may be obtained only by the written authorization of the applicant or his authorized representative and shall be available only during the year of approval. The applicant shall be held responsible for this use. No labels shall be affixed to any item not currently approved by the Department, nor shall the labels be transferred to the possession of any unauthorized person.

Prior to the expiration date of an approval, the applicant shall file the following in order to initiate renewal processing: (1) A completed set of application forms; (2) The required renewal application fee and any supplemental fees if required; and (3) A current sample of the article submitted for approval renewal produced within 90 days of submittal date. Failure to submit any of the foregoing by the expiration date will result in closing of the file.

This approval does not fulfil the requirements for approval which may be required under the Plumbing or Mechanical Code of the City of Los Angeles.

Behrooz Hooriani, P.E.
Electrical Engineer
Electrical Testing Laboratory
2319 Dorris Place
Los Angeles, CA 90031
(323) 224-2168 Tel
(323) 224-2161 Fax

www.LADBS.org



Strong Motion Accelerograph